

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 8 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130295</u>	
Name <u>J Swartz</u>		ID# <u>4314</u>	Date <u>07/01/09</u>
A	Agency <u>SCDL</u>		Phone # <u>269-5592</u>
Instrument Location <u>SCDL</u>			
B	Alco S/N <u>SC07108</u>		Target Value <u>.080</u> High Pressure <u>700 psi</u>
Alco Test Values		<u>.081</u> 1 st Alco	<u>.081</u> 2 nd Alco
Signature <u>[Signature]</u>		<u>[Signature]</u> 7/14/09	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of July 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130295 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

JULY 01, 2009

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI:

L

O.L. #: L

DEPT/AGENCY: ARA1

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .000 ✓

ALCO S/N: SCOTT08

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.37 in

ALCO TARGET .001 09:05

BLANK TEST .000 09:06

INTERNAL STANDARD VERIFIED 09:06

ALCO TV 30.37 in .001 ✓ 09:06

BLANK TEST .000 09:07

SUBJECT SAMPLE .000 09:08

BLANK TEST .000 09:09

ALCO TV 30.37 in .001 ✓ 09:09

BLANK TEST .000 09:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

JULY 01, 2009 ✓

TIME 09:03

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

VENTILATORS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.37 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~